

## Specification

**Incubator- Clean™**

**A5230**

<b>usage</b>	sterile solution
<b>HS-No.:</b>	38220000
<b>Storage:</b>	RT
	<ul style="list-style-type: none"> <li>• Disinfectant solution for incubators and sterile benches.</li> <li>• non-toxic and biodegradable ingredients.</li> <li>• fully compatible with common work surfaces</li> </ul>
	<ul style="list-style-type: none"> <li>• Incubator-Clean™ is a registered trademark of AppliChem GmbH.</li> </ul>
<p><b>Comment</b></p> <p>The problem of contamination in incubators and/or sterile work benches is often a serious one, leading to extensive damage. Incubator-Clean™ solution prevents contamination with and growth of fungi (and spores), bacteria (and spores, including Tuberculosos bacteria), viruses (including HIV and Hepatits B) and mycoplasma. The active ingredients are quaternary benzylammonium compounds. The substance is free of mercury, formaldehyde, phenol or alcohol. It is non corrosive to aluminium, tin-plated iron, chrome, nickel, steel, high grade steel and copper. Furthermore, Incubator-Clean™ is non-toxic and biodegradable.</p> <p><b>Recommended Use:</b> Spray incubators once every 2 weeks. It is not necessary to empty the incubator before spraying, since the active substance is not volatile. Spray sterile benches once a day, or preferably before each laboratory worker begins to use the work area. The drying time is the reaction time!</p>	